

Knowledgebase > Welch Allyn Home > Troubleshooting > Issue 1 - Blood Pressure Device Measurement Issues

Issue 1 - Blood Pressure Device Measurement Issues

Joel Barcillo - 2024-02-09 - Troubleshooting



Issue

The icon above appears on the Blood Pressure Device screen during measurement.

Possible Cause

Too much motion during measurement

Suggested Action

Try to sit still while taking your blood pressure. Motion may result in an inaccurate reading.

Go back to the Patient Help Directory

Go back to Troubleshooting Directory